

ENZYCLEAN IV

Multiple Enzyme Detergent

Enzyclean® IV Multiple Enzyme Detergent is specially formulated for use as a presoak or manual detergent and for use in ultrasonic washers, automatic washing equipment and evacuators. The protease, lipase, amylase and cellulase enzymes in Enzyclean® IV break down organic and inorganic soils, residues, films and deposits and ensure thorough and complete cleaning, leaving instruments moving freely and spot-free.

FEATURES & BENEFITS

- **Four Enzymes.** Protease enzyme effectively breaks down proteins. Lipase enzyme breaks down fats. Amylase enzyme breaks down carbohydrates. Cellulase enzyme breaks down cellulose. All four enzymes work together to prevent biofilm.
- **Rust Inhibitors.** Reduces likelihood of rust, tarnish or corrosion.
- **Rinsing Agents:** Rinses spot & residue-free.
- **Highly Concentrated.** Low cost per use.
- **Neutral pH.** Product safe and compatible with all metals.
- **Low Sudsing.** Compatible with most equipment & pump systems.
- **Lubricating.** Leaves instruments moving freely and prolongs life of instruments.



ORDER INFORMATION ▶ Item Code: 128EME • Size: 1 Gallon (3.8 L) • Case: 4 Units/Case

ENZYCLEAN IV

Multiple Enzyme Detergent



USES Pre-soak • Manual • Ultrasonic • Evacuators • Automatic Washer • Automatic Endoscope Reprocessor

PHYSICAL PROPERTIES Form: Liquid • Color: Blue • Odor: Fresh • pH Concentration: 8.7

TEMPERATURE Usage: 60°-140° F (16°-60° C) • Storage: Cool, dry area

DIRECTIONS

Wear appropriate personal protective equipment based on safe work practices.

PRESOAK, MANUAL AND ULTRASONIC CLEANING

Add 1/8 fl. oz. (3.7 mL) to 2 fl. oz. (59 mL) per gallon (3.8 L) of warm water. Allow to soak for 2-3 minutes or longer. Discard solution after each use.

AUTOMATIC WASHING EQUIPMENT

Add 1/8 fl. oz. (3.7 mL) to 2 fl. oz. (59 mL) per gallon (3.8 L) of water. Recommended temperature range: 90°-140° F (32°-60° C).

EVACUATORS

Add 1/8 fl. oz. (3.7 mL) to 2 fl. oz. (59 mL) per gallon (3.8 L) of water. Suggested maintenance twice weekly.

Dilutions vary depending on temperature, water quality and equipment. Validated cleaning is recommended